

4/25/08

Honeywell

Burdick & Jackson

Honeywell
Burdick & Jackson
1953 South Harvey Street
Muskegon, Michigan 49442-6184

800.368.0050
231.726.3171
FAX 231.728.8226

Certificate of Analysis

Solvent: ACS/HPLC Acetone
Cat: AH010
Lot: CW415

Meets ACS Specifications

Date of Manufacture: 04/24/08

<u>TEST</u>	<u>RESULT</u>
Water by Infrared Absorption	0.43 %
UV Cutoff	328 nm
Refractive Index (20°C)	1.3588
Residue	Not detected (<1 mg/L)
GC Analysis (excluding water)	99.9+ %
Solubility in Water	Passes test
Color (APHA)	Less than 10
Titration Acid	Less than 0.3µeq/g
Titration Base	Less than 0.6µeq/g
Aldehyde (as HCHO)	Less than 0.002 %
Isopropyl Alcohol	Less than 0.05 %
Methanol	Less than 0.05 %
Substances Reducing Permanganate	Passes test

Quality Control Approval



Sample no: 080006039